

Certificate of Conformity

Ex EQUIPMENT

Certificate No.: ANZEx 22.2003X	Current Issue: 0	Date of Issue: 2022-10-18
--	------------------	---------------------------

Applicant: **Acu-Vib Electronics**
Unit 14/22 Hudson Avenue
Castle Hill NSW 2154
Australia

Equipment: Noise Dosimeter type SV 104BIS

Type of Explosion Protection: Intrinsic safety "i"

Explosion Protection Marking: Ex ia I Ma
Ex ia IIC T4 Ga

*This certificate is granted subject to the requirements as set out in
Joint Accreditation System of Australia and New Zealand Publications
ANZEx System Rules 2020 & ANZEx Certified Equipment Scheme Rules 2021*

Signed for and on behalf of issuing body

Name & Position


Geoff Barnier
Principal Engineer - Certification

This certificate is not transferable and remains the property of the issuing body.

The status of this certificate can be confirmed through the database located at www.anzex.com.au

Certificate issued by:

Safety in Mines, Testing and Research Station
2 Robert Smith Street, REDBANK QLD 4301