



**QPS Evaluation Services Inc**  
**Testing, Certification and Field Evaluation Body**  
**Accredited in Canada, the USA, and Internationally**

Page 1 of 1

File  
**LR1356**

**CERTIFICATE OF COMPLIANCE**  
**(ISO TYPE 3 CERTIFICATION SYSTEM)**

Issued to	SVANTEK Sp. z o.o	
Address	Strzyglowska 81 04-872 Warszawa Poland	
Project Number	LR1356-2R1	
Product	Intrinsically Safe Noise Dosimeter	
Model Number	SV104BIS	
Electrical Ratings	Battery Operated Charging parameters: Um = 8 VDC	
Markings	Ex ia IIC T4 Ga Class I, Zone 0, AEx ia IIC T4 Ga Ambient temperature: -10°C to +50°C Degree of protection: IP65	
Applicable Standards	CSA C22.2 No. 60079-0:19 CSA C22.2 No. 60079-11:14	UL 60079-0 7th ed. UL 60079-11 6th ed.
Factory/Manufacturing Location	SVANTEK Sp. z o.o Strzyglowska 81, 04-872 Warszawa, Poland	
Conditions of Certification	<ul style="list-style-type: none"><li>The product shall only be charged in non-hazardous areas by manufacturer's chargers.</li></ul>	

**Statement of Compliance:** The product(s)/equipment identified in this Certificate and described in the Certification Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.

**IMPORTANT NOTE:** In order to maintain the integrity of the QPS Mark(s), certification will be revoked if:

- (1) Compliance to the above-mentioned Standard(s), or those identified in future QPS Standard Update Notice – SUN (QSD 55) is not maintained, or,
- (2) If the product/equipment is modified after certification is granted without prior written consent by QPS



Issued By: Dave Adams, P Eng - Hazardous Locations [Ex Products]

Signature:

Date: September 23, 2024



The SCC and IAS Accreditation Symbols are official symbols of the respective accreditation bodies, used under license.

81 Kelfield St., Unit 8, Toronto, ON M9W 5A3 Tel: 416-241-8857; Fax: 416-241-0682

www.qps.ca

